

TEST STRIP KIT SHIPMENT QC LOG SHEET

Frequency of QC Testing: Test each new Lot #

Date Received	Date Tested	Tested by	PT Test Strips Lot #	Low Control					High Control				
				Lot #	Exp. Date	Expected Range (sec.)*	PT Sec.	Pass (✓)	Lot #	Exp. Date	Expected Range (sec.)*	PT Sec.	Pass (✓)

*Expected range for low and high control strip lot can be found on the test strip kit box.

Note: Control strip lot numbers on the test strip kit label will be different than strip kit lot and patient test strip lot numbers.

- Perform control testing when you receive your shipment as controls have a limited shelf life.
- Perform one high and one low control strip with each new lot of strips received, and record the results on the log.
- If you have multiple lot numbers, perform control tests from each lot received.
- Write "QC DONE" on the top of each test strip kit to note controls were completed on this carton/lot.
- Refer to your user manual for additional information. Call ARKRAY Technical Support at 800-818-8877 if you have questions.